

# **Survey of Antibiotic Test Methods for Raw Milk**

**Commissioned by  
The Dairy Industry Federation**

**DD Muir  
N West  
26 July 1999  
Hannah Research Institute  
AYR KA6 5HL  
Scotland, UK**



**HANNAH  
RESEARCH INSTITUTE**

## Survey of Antibiotic Test Methods - Dairy Industry Federation 1999

### Introduction

Immediately prior to the end of the Milk Marketing schemes the Joint Committee of England & Wales re-issued the "Blue Book" under the title of "Methods for the assessment of Raw Milk Quality". Subsequently the Milk Quality Forum (whose members are the Dairy Industry Federation, National Farmers Union and the UK Federation of Milk Producer Organisations) was set up and agreed to keep this manual up to date.

The Dairy Industry Federation successfully obtained funding from MAFF from its Sector Challenge programme for the development of a raw milk quality programme. Part of this programme was a revision of the "Blue Book". The Milk Quality Forum was aware that the section on antibiotic testing was in particular need of revision, with a large number of antibiotic test kits being offered to milk purchasers. Taking into account the specialised nature of this subject, and the desirability of carrying out practical evaluations of the various tests, the Dairy Industry Federation commissioned the Hannah Research Institute to undertake an examination of the tests available.

Although at the outset of the survey it was anticipated that a single test might be found to serve all the needs of industry it quickly became clear that this objective was unlikely to be met because of the diverse needs of individual dairy companies for antibiotic testing and because of the diversity of the test methods available. These complications precluded the use of an exactly similar protocol for evaluation of the test methods. A compromise was reached where all methods were evaluated in a similar way. The differences being confined to the number of replications of the test at a particular concentration of antibiotic and to the concentration levels of antibiotic tested. The information about each Antibiotic Test Kit is reported in a common format.

This report comprises: (a) Definition of the testing method and rating system, (b) individual reports on all test methods submitted to us and, (c) A brief summary of our findings.

### Definition Of Testing Methods And Rating System

#### Test Samples

Morning milk was collected under near aseptic conditions from healthy cows in the Institute herd that had no history of antibiotic treatment for mastitis or any other condition. The cows were all in mid-lactation and the somatic cell count of the herd was consistently less than 150,000 and the bacterial count index was less than 50 (Bactoscan). The milk was cooled in ice immediately after collection and stored under refrigeration until required. The milk from 5 animals was mixed in equal proportions to give a commingled sample. Spiked samples were prepared by (1) Dilution of antibiotic (0.250g; Penicillin G, sodium salt monohydrate; or Cloxacillin; Sigma Chemical Corp.) in distilled water (100ml), (2) Dilution of the master

solution (1.0 ml) in distilled water (1000 ml) and (3) Addition of the diluted master solution to milk to obtain concentrations of Penicillin G in the range 0-8 mg kg<sup>-1</sup> PPB) and concentrations of Cloxacillin in the range 0-100 mg kg<sup>-1</sup> (PPB). (No correction was applied for the purity of the antibiotic. However, the equivalent concentration of free base is reported based on purity of penicillin G of 93.7% and Cloxacillin of 90.8% derived from data on product label.)

Spiked milk samples were used on the day of preparation and for up to 2 days after collection (stored under refrigeration). As a result, over the testing period for a single antibiotic test kit it was usual for several batches of spiked sample to be used. Care was taken to ensure that the complete range of test concentrations of antibiotic was evaluated for each batch of milk to avoid confounding between batches and the antibiotic concentration. There was no evidence of batch-to-batch variation. During testing, either milk from individual cows was used or the co-mingled milk of 5 animals, depending on the recommendation of the supplier of the test method.

Two antibiotics were selected as the test stimuli - penicillin G and Cloxacillin. The choice of antibiotic was based on the universal acceptance of penicillin G as a benchmark. Cloxacillin was selected as the second test material because of its widespread therapeutic use in the UK. (Note: Procaine penicillin and Cloxacillin rank first and second as active ingredients of the most numerous of licensed proprietary formulations in the UK). A series of concentrations of the two antibiotics in milk were prepared. The choice of levels to be tested was discussed at length with technical experts from the companies supplying the antibiotic tests. Levels were chosen to encompass the expected threshold for detection.

It is important to note that the threshold levels for detection of different antibiotics vary between tests. The test protocol sought to confirm the thresholds for Penicillin G and Cloxacillin. It should also be noted that the reported spectrum of antimicrobials and inhibitory substances detected by the various test methods varies significantly. It was not possible to confirm these claims within the scope of the current study. However, a table reporting the manufacturer specification for test sensitivity of other antimicrobial compounds is included in the report sheets for information.

#### Reporting of test results

The results of the detection tests are reported in tabular form. The recording of a result was according to the instructions of the kit supplier. The exact interpretation of the results is detailed in the individual report forms.

#### Ease of Use Statistic

An important aspect of the method is its ease of use. To avoid subjective assessment, we have devised an 'Ease of Use Statistic' which is objectively based. The potential for deviation from the test protocol is associated with the number (and type) of unit operations (individual

manipulations) which comprise the overall analysis. Thus, by determining the number of independent manipulations in a test, an overall rating for complexity (the inverse of 'ease of use') can be deduced. In the scheme adopted here, low indices indicate a user-friendly format. Each test was rated according to the following criteria detailed in Table 1.

### Testing Schedule

The 17 test kits listed in Table 2 were evaluated. In each and every case a Technical Expert from the supplier visited HRI. On this visit the objectives of the survey were described and, in discussion with the Technical Expert, the number of replications of the test and the concentration levels of the spiked milks were agreed. Hannah personnel were then trained by the Technical Expert. Training was deemed complete when both parties were satisfied that the testing method was fully understood by HRI and a number of tests had been successfully completed. No representatives of the suppliers were present during subsequent evaluations. The results of the individual evaluations were sent to each Technical Expert for comment and the description of the Test, but not the results, were amended in the light of these comments.

### Threshold for Detection of Antibiotic

The ideal performance of an antibiotic test is shown in Figure 1. The response would take the form of a step change where the test would report negative results up to a threshold then, in a step change, switch to reporting positive results. In reality, the response usually takes the form of the actual curve shown in Figure 1. In this case, the probability of reporting a positive result increases from 0 to 1 over a transitional range of antibiotic concentrations. This range and the response curve can be defined by very extensive testing but such an approach was impossible within the resource available for the project. In this study, the threshold value for a test was considered to be the point at which 95% or more of the replicated tests at a particular antibiotic concentration were positive (i.e., 19/20 or 19/30 positive). This approach is conservative but ensures that a sample is not condemned as contaminated unless there is no reasonable doubt that an antimicrobial compound is present.

### Summary of Findings

The test methods examined in this study fell within three categories:

- First, inhibition tests which detect the presence of a wide range of antimicrobial compounds;
- Second, tests specific for a class or limited range of antimicrobial compounds based on an analytical routine with instrumental detection. Such tests require skilled operators.
- Third, special tests designed for use by operators with rudimentary training.

The assignment of test kits into these categories is shown in Table 3.

### Performance - sensitivity to penicillin G and cloxacillin

The sensitivity of all the tests for the detection of penicillin G is reported in Table 4. The Maximum Residue Limit (MRL) for this compound are currently set at 4 ppb (1 ppb = 1 g kg<sup>-1</sup>) within the European Union. Although there was a wide range in sensitivity, with the exception of the Penzyme method, all tests detected penicillin G at or below the MRL. It should be noted that the BetaScreen technique was particularly sensitive to this antimicrobial compound.

The corresponding results for cloxacillin (MRL 30 ppb) are also shown in Table 4. There was greater divergency in the detection limit for cloxacillin - the detection threshold ranged from 5 to 73 ppb. Although not unusually sensitive to penicillin G, the BetaStar and LacTek methods had a detection thresholds for cloxacillin below 7 ppb.

### Performance - ease of use statistic

The ease of use statistics are shown in Table 5. Within each of the categories shown in Table 3, there were significant differences in ease of use. The BetaScreen (EOU -23) and Charm II (EOU - 23/24) tests were more complicated than the Delvo-X-Press (EOU - 10) and LacTek (EOU -11) tests. However, the instrument used for the BetaScreen test also supports the Fluorophos alkaline phosphatase test for checking the efficiency of pasteurisation and the Charm II also may be used in a wider range of applications. The other test methods were either easy to use (BetaStar100, Penzyme 100, Charm AIM; EOU - 9-11) or very easy to use (BetaStar 25, Penzyme 20, Charm MRL, Charm VN, Charm MV, Delvtest P, Delvotest SP and Delvo MCS; EOU - 5-7)

### Performance - length of test

The relative time taken for sample evaluation and the number of samples that may be tested at the same time are shown in Table 6. There are two distinct types of test - those developed for mass screening or where time is not of the essence (Charm AIM, VN & MV, Delvotest P, SP & MCS) and the rapid tests designed for tanker acceptance/rejection at the dairy reception. Of the rapid tests most were complete in 6-12 minutes. However, the BetaScreen and Penzyme tests were completed in 20 minutes.

### Conclusion

The 'consumer' has a choice of test method for detection of antibiotic in milk. Demand can be divided into various choices:

- First, mass screening tests where cost is particularly important or rapid screening tests where time is of the essence.

- Second, tests which have the greatest sensitivity for the widest range of antibiotics or tests that are tuned to detect at the EU maximum residue limits.
- Finally, tests that are highly selective which have a broader spectrum of response.

This survey has clearly revealed that the 'consumer' has a choice within each category. For individual company use this choice must be welcomed because it allows individual companies set their own quality criteria. For example, where dairy companies have a dedicated milk pool, from individual producers under contract or from a producer co-operative, suitable criteria can be mutually agreed and the most appropriate test chosen.

However, when milk is subject to trade outside these close confines the Dairy Industry is faced with a dilemma because it is preferable that equivalent tests are used for application to freely traded milk. The decision on this equivalence and on the tests which meet the needs of the industry are matters for the Milk Quality Forum to decide. Nevertheless, provided the milk meets the EU Maximum Residue Limits - it conforms to the statutory requirements.

#### **Supplementary Information**

At the request of the DIF additional information is provided in the Appendices on:

- (a) The Manufacturers' claimed sensitivity to antibiotics accorded fixed and provisional Maximum Residue Limits by the European Union.
- (b) Prices at June 1999.
- (c) Contact addresses.

*DDM/NW*

*26 July 1999*

**Table 1. Calculation of the Ease of Use Statistic**

| Process                                |   | Code | Minimum | Maximum |
|--|---|------|---------|---------|
| Preparation                            | Set up instrument <sup>a</sup>            | 1    | 0       | 2       |
|  | Set up incubation conditions <sup>b</sup> | 2    | 0       | 1       |
|  | Fluid transfers <sup>c</sup>              | 3    | 0       | n       |
| Analysis                               | Fluid transfers <sup>c</sup>              | 4    | 0       | n       |
|  | Timings <sup>d</sup>                      | 5    | 1       | n       |
|  | Other manipulations <sup>e</sup>          | 6    | 1       | n       |
| Interpretation of results <sup>f</sup> |   | 7    | 0       | 1       |
| Total rating <sup>g</sup>              |   |      | n       | n       |

**Notes on rating criteria:**

<sup>a</sup> 0 = not required; 1 = switch on and allow to warm up, 2 = switch on, allow to warm up, calibrate.

<sup>b</sup> 0 = not required, 1 = set up dry incubator or water bath

<sup>c</sup> 0 = none, n = number of fluid transfers

<sup>d</sup> 0 = none, n = number of separate timed operations

<sup>e</sup> 0 = none, n = number of other manipulations e.g., transfer of test strip to incubator, wash tubes.

<sup>f</sup> 0 = decision printed by instrument, 1 = comparison of band density or interpretation of colour change

<sup>g</sup> Total rating = arithmetic sum of ratings for operations coded 1-7

**Table 2. Summary of Tests Completed at Friday, 23 April 1999**

| Kit                              | Report | Outcome                |
|----------------------------------|--------|------------------------|
| BetaScreen, Advanced Instruments |        |                        |
| BetaStar 25, UCB                 |        |                        |
| BetaStar 100, UCB                |        |                        |
| Penzyme 20, UCB                  |        |                        |
| Penzyme 100 UCB                  |        |                        |
| Charm II                         |        |                        |
| Charm AIM                        |        |                        |
| Charm MRL Beta-lactam test       |        |                        |
| Charm Farm test VN               |        |                        |
| Charm Farm test MV               |        |                        |
| Lactek, Guildhay                 |        |                        |
| Paralux, Guildhay                |        | Withdrawn <sup>a</sup> |
| SNAP, IDEXX                      |        |                        |
| DelvoExpress                     |        | Re-tested <sup>b</sup> |
| Delvo P                          |        |                        |
| Delvo SP                         |        |                        |
| Delvo MCS                        |        |                        |

<sup>a</sup> The instrument supplied was a prototype and did not perform according to expectations of the manufacturer. The evaluation was therefore abandoned at the request of the supplier. It is anticipated that the test be re-submitted for evaluation later in the year.

<sup>b</sup> The system did not perform according to expectations in our laboratory, despite re-testing. Previous independent tests have revealed no problem. In collaboration with the manufacturer we are currently investigating the problem.



**Table 3 Assignment of Test Methods into Categories**

| <b>Category 1</b> | <b>Category 2</b> | <b>Category 3</b> |
|-------------------|-------------------|-------------------|
| Charm AIM         | BetaScreen        | BetaStar 25/100   |
| Charm Farm VN     | Charm II          | Charm MRL         |
| Charm Farm MV     | DelvoExpress      | Penzyme 20/100    |
| Delvotest MCS     | Lactek            | SNAP              |
| Delvotest P       |                   |                   |
| Delvotest SP      |                   |                   |

Table 4. Sensitivity of test methods to penicillin G and cloxacillin ( g kg<sup>-1</sup>)

| Test         | Pen. G | Cloxacillin | Test      | Pen. G | Cloxacillin | Test          | Pen. G | Cloxacillin |
|--------------|--------|-------------|-----------|--------|-------------|---------------|--------|-------------|
| BetaScreen   | 0.9    | 20          | Charm AIM | 2.8    | 32          | Delvo-X-Press | 4.0    | 60          |
| BetaStar 25  | 2.8    | 5.4         | Charm MRL | 2.8    | 27          | Delvotest P   | 2.8    | 36          |
| BetaStar 100 | 2.8    | 5.4         | Charm VN  | 2.8    | 36          | Delvo SP      | 2.8    | 18          |
| Penzyme 20   | 4.7    | 73          | Charm MV  | 2.8    | 27          | Delvo MCS     | 1.9    | 14          |
| Penzyme 100  | 5.6    | 73          | LacTek    | 3.3    | 6.8         |               |        |             |
| Charm II     | 2.8    | 18          | SNAP      | 3.7    | 54          |               |        |             |

Maximum residue limits set by EU are: penicillin G = 4.0 g kg<sup>-1</sup>, cloxacillin = 30.0 g kg<sup>-1</sup>

Table 5. Ease of use statistics (EOU; lower values indicate easiest to use).

| Test         | EOU   | Test      | EOU | Test          | EOU |
|--------------|-------|-----------|-----|---------------|-----|
| BetaScreen   | 23    | Charm AIM | 10  | Delvo-X-Press | 10  |
| BetaStar 25  | 6     | Charm MRL | 5   | Delvotest P   | 5   |
| BetaStar 100 | 9/10  | Charm VN  | 6   | Delvo SP      | 5   |
| Penzyme 20   | 6     | Charm MV  | 7   | Delvo MCS     | 5   |
| Penzyme 100  | 10    | LacTek    | 11  |               |     |
| Charm II     | 23/24 | SNAP      | 7   |               |     |

Table 6. Relative time taken for sample evaluation and number of samples which may be tested simultaneously.

| Test         | No. samples | Time, min. | Test          | No. samples | Time, min. |
|--------------|-------------|------------|---------------|-------------|------------|
| BetaScreen   | 5           | 20         | Charm AIM     | 96          | 240        |
| BetaStar 25  | 4           | 6          | Charm MRL     | 10          | 10         |
| BetaStar 100 | 8           | 6          | Charm VN      | 150         | 210        |
| Penzyme 20   | 8           | 20         | Charm MV      | 72          | 180        |
| Penzyme 100  | 8           | 20         | LacTek        | 5           | 10         |
| Charm II     | 5           | 12         | SNAP          | 6           | 10         |
|              |             |            | Delvo-X-Press | 5           | 8          |
|              |             |            | Delvotest P   | 40          | 150        |
|              |             |            | Delvo SP      | 40          | 210        |
|              |             |            | Delvo MCS     | 150         | 150        |

Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |   |
|---------------------------------|---|
| <b>Name:</b>                    | Beta-Screen EU (Fluorophos®)  |
| <b>Supplier</b>                 | Advanced Instruments Inc., Norwood, MA, USA   |
| <b>UK Agent</b>                 | QuadraChem, Riverside, Forest Row Business Park, Forest Row, East Sussex RH18 5DW   |
| <b>Expert Technical Contact</b> | Jeffrey P. Guilbert ( <a href="mailto:jeffe@aitests.com">jeffe@aitests.com</a> )  |
| <b>Recommended application</b>  | Qualitative test for beta-lactam antibiotic residues in commingled bovine milk including: penicillin G, amoxicillin, ampicillin, cloxacillin, dicloxacillin and oxacillin. Applicable to pasteurised milk.  |
| <b>Principle of test</b>        | BetaScreen is a competitive enzyme-linked immunoassay. Milk and enzyme-conjugate are added to a tube coated with an antibody to beta-lactam antibiotic. If no antibiotic is present in the milk, the binding sites are complexed only with the enzyme-conjugate. After washing, when a substrate is added, the enzyme releases a fluorescent material that is detected by a fluorimeter. When antibiotic is present in milk, it competes with enzyme conjugate for the binding sites on the coated tube. As a result, a mixture of enzyme-conjugate and antibiotic are bound to the tube. Therefore, after washing and addition of substrate less fluorescent material is released. The decrease in fluorescence is related to the concentration of antibiotic. |
| <b>Cross-reactivity</b>         | BetaScreen does not cross react at a concentration of 100 parts per billion with ceftriaxone, cephalosporin-C, cephapirin, chlorothiazide, chlortetracycline, dexamethasone, doxycycline, dipyrone, erythromycin, furosemide, gentamicin, ivermectin, neomycin, novobiocin, oxytetracycline, oxytocin, para-aminobenzoic acid, phenylbutasone, pirlimycin, streptomycin, sulphadiazine, sulphadimethoxine, sulphamethazine, sulphamonomethoxine, sulphapyridine, sulphathiazole, tetracycline, thiabendazole, tilmycosin and trichlormethiazide.  |
| <b>Interference</b>             | Sensitivity and selectivity are not adversely affected by somatic cell numbers up to 800,000 cells per ml.  |
| <b>Apparatus supplied</b>       | In start up kit:- Tray organiser, timer, graduated cylinder (50 ml), wash bottle (500ml), pipette (500 µl) and disposable tips, foam assay tube holder.   |
| <b>Extra apparatus required</b> | Fluorimeter (Fluorophos Test System), vortex mixer and cuvettes.  |
| <b>Reagents supplied</b>        | Beta Standard and diluent, Beta-conjugate, Fluorophos substrate, substrate buffer, stop solution, wash concentrate, assay tubes.  |
| <b>Extra reagents required</b>  | Distilled or deionised water  |
| <b>Time (set-up)</b>            | 30-40 minutes   |
| <b>Time per test</b>            | 20 minutes (5 samples are tested together; time per test 3 minutes)   |
| <b>Operator skill level</b>     | Training required by an agent of Advanced Instruments Inc.; previous experience in analytical procedures is helpful.  |
| <b>Instructions</b>             | Detailed and unambiguous.   |
| <b>Safety advice</b>            | Hazard warnings are detailed on reagent bottles.  |
| <b>Reporting method</b>         | Instrument prints out 'positive' or 'negative' result   |
| <b>Ease of use</b>              | Requires careful adherence to instructions (Rating = 23).   |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 20         | 0            | 0     | 0      | 20         | 0            |
| 0.25   | 0.23    | 20         | 3            | 10    | 9.1    | 20         | 19           |
| 0.5    | 0.46    | 20         | 8            | 15    | 13.7   | 20         | 17           |
| 1.0    | 0.94    | 20         | 20           | 20    | 18.2   | 20         | 15           |
| 2.0    | 1.9     | 20         | 20           | 22.5  | 20.4   | 20         | 20           |
| 3.0    | 2.8     | 20         | 20           | 30    | 27.2   | 20         | 20           |
|        |         |            |              | 37.5  | 34.1   | 20         | 20           |

Positive attribution when ratio of Test/Standard readings 1.0

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic    | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|---------------|---------------------------------------|
| Penicillin G  | 1                                     |
| Amoxicillin   | 10                                    |
| Ampicillin    | 3                                     |
| Cloxacillin   | 20                                    |
| Dicloxacillin | 30                                    |
| Oxacillin     | 20                                    |
| nafcillin     | 10                                    |

*Additional Comments*

The fluorimeter used to measure the extent of reaction is common to Fluorophos Test systems for alkaline phosphatase (used to detect efficient pasteurisation of milk) and a test under development (acid phosphatase) for confirmation of the efficient pasteurisation of meat.

*DDM/NW March 12, 1999 (Final revision)*

## Test Report

Test method: Beta s.t.a.r 25  
Beta s.t.a.r 100

UK supplier: Axient Laboratories

**Dairy Industry Federation Survey - Report on Antibiotic Test Kit**

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Beta s.t.a.r. 100  |
| <b>Supplier</b>                 | UCB Bioproducts, Braine-L'Alleud, Belgium  |
| <b>UK Agent</b>                 | Axient, Axient Laboratories, Unit 20, Goldthorpe, Rotherham S63 9BL  |
| <b>Expert Technical Contact</b> | Jacques Degelaen ( <a href="mailto:jacques.degelaen@ucb-group.com">jacques.degelaen@ucb-group.com</a> )  |
| <b>Recommended application</b>  | Qualitative test for beta-lactam antibiotics, including penicillin G, ampicillin, amoxicillin, nafcillin, cloxacillin, oxacillin, dicloxacillin, cephapirin, cephalonium, cefoperazone and ceftiofur. Beta s.t.a.r 100 is designed for screening milk in any laboratory where an answer is needed with respect to contamination of milk by beta lactam antibiotics.  |
| <b>Principle of test</b>        | The test involves a specific beta-lactam receptor linked to gold particles. An initial incubation of a specific amount of receptor with milk containing antibiotics results in a complex between the receptor and the antibiotic. Second, the solution is transferred onto an immunochromatographic medium. The first band captures receptors that have not complexed with antibiotic. The second band serves as a reference band. |
| <b>Cross-reactivity</b>         | Not available at time of test  |
| <b>Interference</b>             | Operator should not be under medical treatment by beta-lactam antibiotics.   |
| <b>Apparatus supplied</b>       | None.  |
| <b>Extra apparatus required</b> | Dry incubator or water bath at $47^{\circ}\pm 0.5C$ , micropipettes (25 & 100 $\mu$ l), pipette to deliver 1350 $\mu$ l.   |
| <b>Reagents supplied</b>        | 2 vials of receptor, 4 x 25 dipsticks (receptor can be stored in deep freeze for more than a month)  |
| <b>Extra reagents required</b>  | Deionised water  |
| <b>Time (set-up)</b>            | Warm-up time of dry incubator, preparation of solutions  |
| <b>Time per test</b>            | 6 minutes (8 samples may be tested in one batch)   |
| <b>Operator skill level</b>     | Minimal  |
| <b>Instructions</b>             | Detailed and unambiguous.  |
| <b>Safety advice</b>            | No specific hazard   |
| <b>Reporting method</b>         | After development, the test strips normally show two distinct bands, an upper (reference) and a lower. If the lower band is denser than the reference, the sample is negative; if the bands are similar in density then a positive with caution (+/-) result should be applied; if the lower band is less dense than the reference (+) or absent (++) then the sample is positive.   |
| <b>Ease of use</b>              | Relatively easy to use (rating = 9/10).  |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 20         | 0            | 0     | 0      | 20         | 0            |
| 2      | 1.8     | 20         | 0            | 2     | 1.8    | 20         | 0            |
| 2.5    | 2.3     | 20         | 1            | 3     | 2.7    | 20         | 0            |
| 3      | 2.8     | 20         | 20 (-/+)     | 6     | 5.4    | 20         | 0            |
| 4      | 3.8     | 20         | 20 (+)       | 9     | 8.2    | 20         | 20 (+)       |
| 5      | 4.7     | 20         | 20 (++)      | 12    | 10.9   | 20         | 20 (+)       |
| 6      | 5.6     | 20         | 20 (++)      |       |        | 20         | 20 (++)      |
| 8      | 7.50    | 20         | 20 (++)      |       |        |            |              |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic    | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|---------------|---------------------------------------|
| Penicillin G  | 2-4                                   |
| Amoxicillin   | 2-4                                   |
| Ampicillin    | 2-5                                   |
| Nafcillin     | 8-20                                  |
| Oxacillin     | 5-10                                  |
| Cloxacillin   | 5-10                                  |
| Dicloxacillin | 5-10                                  |
| Cephapirin    | 8-16                                  |
| Cephalonium   | 7.5-15                                |
| Cefoperazone  | 5-8                                   |
| Ceftiofur     | 75-150                                |

*Additional Comments*

Reacted strips can be stored as a permanent record.

DDM/IV March 12, 1999 (Final revision)

**Dairy Industry Federation Survey - Report on Antibiotic Test Kit**

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Beta s.t.a.r. 25   |
| <b>Supplier</b>                 | UCB Bioproducts, Braine-L'Alleud, Belgium  |
| <b>UK Agent</b>                 | Axient, Axient Laboratories, Unit 20, Goldthorpe, Rotherham S63 9BL  |
| <b>Expert Technical Contact</b> | Jacques Degelaen ( <a href="mailto:jacques.degelaen@ucb-group.com">jacques.degelaen@ucb-group.com</a> )  |
| <b>Recommended application</b>  | Qualitative test for beta-lactam antibiotics, including penicillin G, ampicillin, amoxicillin, nafcillin, cloxacillin, oxacillin, dicloxacillin, cephapirin, cephalonium, cefoperazone and ceftiofur. Beta s.t.a.r is designed for screening milk in any situation where an answer is needed with respect to contamination of milk by beta lactam antibiotics.   |
| <b>Principle of test</b>        | The test involves a specific beta-lactam receptor linked to gold particles. An initial incubation of a specific amount of receptor with milk containing antibiotics results in a complex between the receptor and the antibiotic. Second, the solution is transferred onto an immunochromatographic medium. The first band captures receptors that have not complexed with antibiotic. The second band serves as a reference band. |
| <b>Cross-reactivity</b>         | Not available at time of testing   |
| <b>Interference</b>             | The operator should not be under medical treatment using beta-lactam antibiotics.  |
| <b>Apparatus supplied</b>       | Spring syringe and 25 disposable tips.   |
| <b>Extra apparatus required</b> | Dry incubator or water bath at $47^{\circ} \pm 0.5C$   |
| <b>Reagents supplied</b>        | 25 individual vials of receptor, 25 dipsticks  |
| <b>Extra reagents required</b>  | None   |
| <b>Time (set-up)</b>            | Warm-up time of dry incubator  |
| <b>Time per test</b>            | 6 minutes (4 samples may be tested at one time)  |
| <b>Operator skill level</b>     | Minimal  |
| <b>Instructions</b>             | Detailed and unambiguous.  |
| <b>Safety advice</b>            | No specific hazards  |
| <b>Reporting method</b>         | After development, the test strips normally show two distinct bands, an upper (reference) and a lower. If the lower band is denser than the reference, the sample is negative; if the bands are similar in density then a positive with caution (+/-) result should be applied; if the lower band is less dense than the reference (+) or absent (++) then the sample is positive.   |
| <b>Ease of use</b>              | Very easy to use (Rating = 6)  |



*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 20         | 0            | 0     | 0      | 20         | 0            |
| 2      | 1.8     | 20         | 0            | 2     | 1.8    | 20         | 0            |
| 2.5    | 2.3     | 20         | 0            | 3     | 2.7    | 20         | 0            |
| 3      | 2.8     | 20         | 20 (+)       | 6     | 5.4    | 20         | 20 (+)       |
| 4      | 3.8     | 20         | 20 (++)      | 9     | 8.2    | 20         | 20 (+)       |
| 5      | 4.7     | 20         | 20 (++)      | 12    | 10.9   | 20         | 20 (++)      |
| 6      | 5.6     | 20         | 20 (++)      |       |        |            |              |
| 8      | 7.50    | 20         | 20 (++)      |       |        |            |              |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic    | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|---------------|---------------------------------------|
| Penicillin G  | 2-4                                   |
| Amoxicillin   | 2-4                                   |
| Ampicillin    | 2-5                                   |
| Nafcillin     | 8-20                                  |
| Oxacillin     | 5-10                                  |
| Cloxacillin   | 5-10                                  |
| Dicloxacillin | 5-10                                  |
| Cephapirin    | 8-16                                  |
| Cephalonium   | 7.5-15                                |
| Cefoperazone  | 5-8                                   |
| Ceftiofur     | 75-150                                |

*Additional Comments*

Reacted strips can be stored as a permanent record.

*DDM/NW March 12, 1999 (Final revision)*

**Dairy Industry Federation Survey - Report on Antibiotic Test Kit**

|                                 |   |
|---------------------------------|---|
| <b>Name:</b>                    | <b>CharmII® Beta-lactam Test for Maximum Residue Limits</b>   |
| <b>Supplier</b>                 | CHARM Sciences Inc., 36 Franklin Street, Malden MA USA  |
| <b>UK Agent</b>                 | FOSS UK Ltd, Parkway House, Station Road, Didcot Oxon OX11 7NN  |
| <b>Expert Technical Contact</b> | Bob Salter, VP.Regulatory Affairs, E.mail to charm1@world.std.com   |
| <b>Recommended application</b>  | The Charm II beta-lactam Test detects beta-lactams at or below the maximum residue limits in raw, comingled bovine milk. It is designed for use by milk, intake, laboratory, field and regulatory personnel.  |
| <b>Principle of test</b>        | The test uses bacteria with specific receptor sites that bind all beta-lactam drugs. The bacteria are added to a milk sample together with a minute amount of <sup>14</sup> C labelled penicillin G. Any beta-lactam already in the milk competes with this labelled penecillin G. The amount of <sup>14</sup> C labelled penicillin G that binds to the receptor sites is measured compared to a previously determined control point. The greater the amount of <sup>14</sup> C labelled penicillin G measured, the lower the beta-lactam concentration in the sample. |
| <b>Cross-reactivity</b>         | No interference (at 100ppb) from sulfadiazine, sulfanilamide, sulfathiazole, sulfamethazine, sulfapyridine, sulfadimethoxine, tetracycline, oxytetracycline, chlortetracycline, doxycycline, gentamicin, neomycin, streptomycin, ivermectin, erythromycin, novobiocin, furosemide, trichlormethiazide, chlorothiazide, oxytocin, phenylbutazone, dexamethasone, PABA and dipyrone.  |
| <b>Interference</b>             | Intra-family cross reactivity with cefadroxil (15ppb), cefotaxime (4ppb), cephalixin (15ppb), cephradine (15ppb), hetacillin (5ppb), piperacillin (8ppb) and ticarcillin (35ppb)  |
| <b>Apparatus supplied</b>       | None  |
| <b>Extra apparatus required</b> | Charm scintillation counter, test tubes, caps for test tubes, cotton swabs, pipette tips (5 ml & 1 ml)  |
| <b>Reagents supplied</b>        | Beta-lactam tablet reagents, operators manual, MRL standard , zero control standard, standards manual, typical counts sheet   |
| <b>Extra reagents required</b>  | Deionised water, scintillation fluid  |
| <b>Time (set-up)</b>            | 1 hour  |
| <b>Time per test</b>            | 12 minutes (a control and 5 tests may be run at the same time)  |
| <b>Operator skill level</b>     | Moderate; training required; videotape available  |
| <b>Instructions</b>             | Concise and clear   |
| <b>Safety advice</b>            | Explicit  |
| <b>Reporting method</b>         | Positive or negative in relation to a pre-determined control point  |
| <b>Ease of use</b>              | Requires training and attention to detail (Ease of use = 23/24)   |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 1            |
| 1.0    | 0.9     | 30         | 6            | 15    | 13.7   | 30         | 25           |
| 2.0    | 1.9     | 30         | 23           | 20    | 18.2   | 30         | 29           |
| 3.0    | 2.8     | 30         | 30           | 25    | 22.8   | 30         | 30           |
| 4      | 3.8     | 30         | 30           | 30    | 27.3   | 30         | 30           |
| 6      | 5.6     | 30         | 30           | 35    | 31.9   | 30         | 30           |
| 8      | 7.5     | 30         | 30           | 40    | 36.4   | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic    | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|---------------|---------------------------------------|
| Penicillin G  | 2                                     |
| Cefazolin     | 15                                    |
| Ceftiofur     | 40                                    |
| Cefquinome    | 20                                    |
| Cephapirin    | 3                                     |
| Amoxicillin   | 5                                     |
| Ampicillin    | 4                                     |
| Cloxacillin   | 30                                    |
| Dicloxacillin | 20                                    |
| Oxacillin     | 30                                    |
| Penethamate   | 2                                     |
| Nafcillin     | 30                                    |

*Additional Comments*

The contents of the  $^{14}\text{C}$  labelled penicillin G supplied with the Charm Test are sufficiently low that they are exempt from Nuclear Regulatory Commission regulations. Each tablet contains less than 0.15 kilobecquerels of  $^{14}\text{C}$  labelled penicillin G. Solid waste may be disposed of without additional precautions. Liquid waste contaminated with  $^{14}\text{C}$  labelled penicillin G can be flushed with water down the drain. However, the liquid waste contains scintillant which may be collected and disposed of with waste solvent.

*DDM/NIV March 12, 1999 (Third revision)*

### Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Charm AIM-96™ Antimicrobial Inhibition Monitor for 96 Multiwell Plates   |
| <b>Supplier</b>                 | CHARM Sciences Inc., 36 Franklin Street, Malden MA USA   |
| <b>UK Agent</b>                 | FOSS UK Ltd, Parkway House, Station Road, Didcot Oxon OX11 7NN   |
| <b>Expert Technical Contact</b> | Bob Salter, VP.Regulatory Affairs, E.mail to charml@world.std.com  |
| <b>Recommended application</b>  | The Charm AIM-96 is a microbial inhibition assay designed for high volume, broad spectrum screening of raw, pasteurised, homogenised or skim milk. Cream may be run if pre-treated.          |
| <b>Principle of test</b>        | The test detects inhibition of growth of bacteria. A dye is present in the growth medium which changes colour when bacterial growth occurs. If growth is inhibited, no colour change occurs. |
| <b>Cross-reactivity</b>         | Test detects a very wide range of microbial inhibitors   |
| <b>Interference</b>             |  |
| <b>Apparatus supplied</b>       | Multiwell plates, sealing strip, plate lid   |
| <b>Extra apparatus required</b> | AIM-96 Incubator, automatic pipette (50 l), Repitter or octapette with reservoir; reference colour strip   |
| <b>Reagents supplied</b>        | Lyophilised medium, B.stearothermophilus spore tablets   |
| <b>Extra reagents required</b>  | Deionised water, antimicrobial drug free milk  |
| <b>Time (set-up)</b>            | 30 minutes   |
| <b>Time per test</b>            | 3-4 hours, 90 test samples plus 3 negative and three positive controls may be conveniently run at the same time  |
| <b>Operator skill level</b>     | Moderate; some skill in pipetting required   |
| <b>Instructions</b>             | Concise and clear  |
| <b>Safety advice</b>            | No specific hazard   |
| <b>Reporting method</b>         | Positive or negative based on reference to colour chart  |
| <b>Ease of use</b>              | Requires minimal training (Ease of use = 10)   |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 0            |
| 1.0    | 0.9     | 30         | 0            | 20    | 18     | 30         | 9            |
| 2.0    | 1.9     | 30         | 8            | 25    | 23     | 30         | 14           |
| 3.0    | 2.8     | 30         | 29           | 30    | 27     | 30         | 19           |
| 4      | 3.8     | 30         | 30           | 35    | 32     | 30         | 30           |
| 6      | 5.6     | 30         | 30           | 40    | 36     | 30         | 30           |
| 8      | 7.5     | 30         | 30           | 50    | 45     | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

(Antimicrobial drugs listed are typical of their respective families)

| Antibiotic      | Sensitivity* ( $\mu\text{g kg}^{-1}$ ) |
|-----------------|--|
| Penicillin G    | 3-5                                    |
| Sulfamethazine  | 10-50                                  |
| Gentamicin      | 30-100                                 |
| Oxytetracycline | 150-300                                |
| Tylosin         | 40-60                                  |

- \* First level = colour 4, Second level = colour 5

*Additional Comments*

None

DDM/NW March 12, 1999 (Third revision)

### Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Charm MRL™ Beta Lactam Test  |
| <b>Supplier</b>                 | CHARM Sciences Inc., 36 Franklin Street, Malden MA USA   |
| <b>UK Agent</b>                 | FOSS UK Ltd, Parkway House, Station Road, Didcot Oxon OX11 7NN   |
| <b>Expert Technical Contact</b> | Bob Salter, VP.Regulatory Affairs, E.mail to charm1@world.std.com  |
| <b>Recommended application</b>  | The Charm MRL™ Test is a rapid receptor assay designed for detection of beta lactam drugs in raw, commingled bovine milk at or near the European Union maximum residue limit. The test is designed for use by milk intake, laboratory, field and regulatory personnel.   |
| <b>Principle of test</b>        | The test uses receptors that bind to beta lactam drugs. As the milk flows through the test strip a line forms the test position when no beta lactam is present in the milk. When beta lactams are present in the sample this line is less dense or absent. The test line is compared to a control line designed to discriminate beta lactams close to the maximum residue limit. The control and test lines were compared using a CHARM LUM-T with imager. |
| <b>Cross-reactivity</b>         | The following drugs at 100ppb show no interference: sulfadiazine, sulfanilamide, sulfathiazole, sulfamethazine, sulfapyridine, sulfadimethoxine, tetracycline, oxytetracycline, chlortetracycline, doxycycline, gentamicin, neomycin, streptomycin, ivermectin, erythromycin, novobiocin, furosemide, trichlomethiazide, chlorothiazide, oxytocin, phenylbutazone, dexamethasone, PABA and dipyrone.   |
| <b>Interference</b>             | Other beta-lactam drugs are detected.  |
| <b>Apparatus supplied</b>       | None   |
| <b>Extra apparatus required</b> | Strip incubator, 300 μl micropipette and disposable tips; optional Imager  |
| <b>Reagents supplied</b>        | MRL Beta lactam Test Strips, operators manual; 4ppb Penicillin G standard  |
| <b>Extra reagents required</b>  | Antimicrobial drug free milk   |
| <b>Time (set-up)</b>            | 20 minutes to allow heater block to equilibrate  |
| <b>Time per test</b>            | <10 minutes, 10 test samples plus negative and positive controls may be conveniently run at the same time if 3 x 4 place incubators are used.  |
| <b>Operator skill level</b>     | Basic; some skill in pipetting required  |
| <b>Instructions</b>             | Concise and clear  |
| <b>Safety advice</b>            | No unusual hazard.   |
| <b>Reporting method</b>         | Positive or negative reported by strip reader  |
| <b>Ease of use</b>              | Requires minimal training (Ease of use = 5)  |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 0            |
| 1.0    | 0.9     | 30         | 12           | 10    | 9.1    | 30         | 6            |
| 2.0    | 1.9     | 30         | 23           | 15    | 13.7   | 30         | 10           |
| 3.0    | 2.8     | 30         | 29           | 20    | 18.2   | 30         | 20           |
| 4      | 3.8     | 30         | 30           | 25    | 22.8   | 30         | 23           |
| 6      | 5.6     | 30         | 30           | 30    | 27.3   | 30         | 30           |
| 8      | 7.5     | 30         | 30           | 35    | 31.9   | 30         | 30           |
|        |         |            |              | 40    | 36.4   | 30         | 30           |
|        |         |            |              | 50    | 45.5   | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic   | Sensitivity ( $\mu\text{g kg}^{-1}$ )* |
|--------------|--|
| Penicillin G | 3                                      |
| Amoxicillin  | 4                                      |
| Ampicillin   | 4                                      |
| Ceftiofur    | 100                                    |
| Cephapirin   | 10                                     |
| Cloxacillin  | 30                                     |
| Oxacillin    | 50                                     |

\*100% positive using Imager

*Additional Comments*

None

*DDM/NW March 24, 1999 (Final revision)*

Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Charm Farm Test-Vial for Beta lactams, sulfonamides and other antimicrobial drugs  |
| <b>Supplier</b>                 | CHARM Sciences Inc., 36 Franklin Street, Malden MA USA   |
| <b>UK Agent</b>                 | FOSS UK Ltd, Parkway House, Station Road, Didcot Oxon OX11 7NN   |
| <b>Expert Technical Contact</b> | Bob Salter, VP.Regulatory Affairs, E.mail to charm1@world.std.com  |
| <b>Recommended application</b>  | The Charm Farm Test-Vial is a microbial inhibition assay designed for detection of beta lactam drugs in raw, commingled milk at or near the European Union maximum residue limit.  |
| <b>Principle of test</b>        | The test detects inhibition of growth of bacteria. A dye is present in the growth medium which changes colour when bacterial growth occurs. If growth is inhibited, no colour change occurs.   |
| <b>Cross-reactivity</b>         | Test detects a very wide range of microbial inhibitors. The following drugs at 100ppb show no interference: ivermectin, novobiocin, furosemide, trichloromethiazide, chlorothiazide, oxytocin, phenylbutazone, dexamethasone and dipyron |
| <b>Interference</b>             | Lacteal secretions from mastitic animals have been reported to contain natural inhibitors that cause positives. No interference from somatic cell ( $10^6$ SCC) or bacteria ( $3 \times 10^5$ cfu/ml)                                    |
| <b>Apparatus supplied</b>       | None   |
| <b>Extra apparatus required</b> | Inctronic incubator or mini incubator or water bath or Delvo incubator, water bath rack, reference colour strip and timer  |
| <b>Reagents supplied</b>        | B.stearothermophilus spore tablets, Charm-VN test vials, 200 $\mu$ l micropipette and disposable tips; 4ppb penicillin G standard  |
| <b>Extra reagents required</b>  | Antimicrobial drug free milk   |
| <b>Time (set-up)</b>            | 20 minutes to allow heater block to equilibrate  |
| <b>Time per test</b>            | <3½ hours, 150 test samples plus negative and positive controls may be conveniently run at the same time   |
| <b>Operator skill level</b>     | Basic; some skill in pipetting required  |
| <b>Instructions</b>             | Concise and clear  |
| <b>Safety advice</b>            | No unusual hazard.   |
| <b>Reporting method</b>         | Positive or negative based on reference to colour chart  |
| <b>Ease of use</b>              | Requires minimal training (Ease of use = 6)  |



*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 0            |
| 1.0    | 0.9     | 30         | 0            | 10    | 9.1    | 30         | 1            |
| 2.0    | 1.9     | 30         | 2            | 15    | 13.7   | 30         | 2            |
| 3.0    | 2.8     | 30         | 30           | 20    | 18.2   | 30         | 2            |
| 4      | 3.8     | 30         | 30           | 25    | 22.8   | 30         | 3            |
| 6      | 5.6     | 30         | 30           | 30    | 27.3   | 30         | 16           |
| 8      | 7.5     | 30         | 30           | 35    | 31.9   | 30         | 28           |
|        |         |            |              | 40    | 36.4   | 30         | 30           |
|        |         |            |              | 50    | 45.5   | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic      | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|-----------------|---------------------------------------|
| Penicillin G    | 3-4                                   |
| Amoxicillin     | 4                                     |
| Ampicillin      | 4                                     |
| Ceftiofur       | 50                                    |
| Cephapirin      | 10                                    |
| Sulfamethazine  | 50-100                                |
| Gentamicin      | 100-300                               |
| Oxytetracycline | 100-150                               |
| Tylosin         | 40-50                                 |

*Additional Comments*

None

*DDM/NW March 24, 1999 (Final revision)*

**Dairy Industry Federation Survey - Report on Antibiotic Test Kit**

|                                 |   |
|---------------------------------|---|
| <b>Name:</b>                    | Charm Farm Test Mini-Vial for Beta lactams, sulfonamides and other antimicrobial drugs  |
| <b>Supplier</b>                 | CHARM Sciences Inc., 36 Franklin Street, Malden MA USA  |
| <b>UK Agent</b>                 | FOSS UK Ltd, Parkway House, Station Road, Didcot Oxon OX11 7NN  |
| <b>Expert Technical Contact</b> | Bob Salter, VP.Regulatory Affairs, E.mail to charml@world.std.com   |
| <b>Recommended application</b>  | The Charm Farm Test Mini-Vial is a microbial inhibition assay designed for detection of beta lactam drugs in raw, commingled milk at or near the European Union maximum residue limit.  |
| <b>Principle of test</b>        | The test detects inhibition of growth of bacteria. A dye is present in the growth medium which changes colour when bacterial growth occurs. If growth is inhibited, no colour change occurs.  |
| <b>Cross-reactivity</b>         | Test detects a very wide range of microbial inhibitors. The following drugs at 100ppb show no interference: ivermectin, novobiocin, furosemide, trichlormethiazide, chlorothiazide, oxytocin, phenylbutazone, dexamethasone and dipyrrone |
| <b>Interference</b>             | Lacteal secretions from mastitic animals have been reported to contain natural inhibitors that cause positives. No interference from somatic cell ( $10^6$ SCC) or bacteria ( $3 \times 10^5$ cfu/ml)                                     |
| <b>Apparatus supplied</b>       | None  |
| <b>Extra apparatus required</b> | Air incubator or water bath or Delvo incubator with mini-vial adaptors, water bath rack, tweezers, 0.1 ml micropipette and disposable tips, reference colour strip and timer  |
| <b>Reagents supplied</b>        | Media tablets, Charm-MV test vials (96 per pack), zero control milk powder, 4ppb penicillin G standard, 12x8 well strip sealing tape  |
| <b>Extra reagents required</b>  | None  |
| <b>Time (set-up)</b>            | 20 minutes to allow heater block or incubator to equilibrate  |
| <b>Time per test</b>            | < 3 hours, 72 test samples plus negative and positive controls may be conveniently run at the same time   |
| <b>Operator skill level</b>     | Basic; some skill in pipetting required   |
| <b>Instructions</b>             | Concise and clear   |
| <b>Safety advice</b>            | No unusual hazard.  |
| <b>Reporting method</b>         | Positive or negative based on reference to colour chart   |
| <b>Ease of use</b>              | Requires minimal training (Ease of use = 7)   |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 0            |
| 1.0    | 0.9     | 30         | 0            | 10    | 9.1    | 30         | 2            |
| 2.0    | 1.9     | 30         | 6            | 15    | 13.7   | 30         | 4            |
| 3.0    | 2.8     | 30         | 30           | 20    | 18.2   | 30         | 4            |
| 4      | 3.8     | 30         | 30           | 25    | 22.8   | 30         | 27           |
| 6      | 5.6     | 30         | 30           | 30    | 27.3   | 30         | 30           |
| 8      | 7.5     | 30         | 30           | 35    | 31.9   | 30         | 30           |
|        |         |            |              | 40    | 36.4   | 30         | 30           |
|        |         |            |              | 50    | 45.5   | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic      | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|-----------------|---------------------------------------|
| Penicillin G    | 3-4                                   |
| Amoxicillin     | 4                                     |
| Ampicillin      | 4                                     |
| Ceftiofur       | 50                                    |
| Cephapirin      | 10                                    |
| Sulfamethazine  | 50-100                                |
| Gentamicin      | 300-500                               |
| Oxytetracycline | 100-150                               |
| Tylosin         | 40-50                                 |

*Additional Comments*

None

*DDM/NW March 24, 1999 (Final revision)*

Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Delvotest® MCS   |
| <b>Supplier</b>                 | Gist-Brocades B.V. PO Box 1, 2600 MA Delft, The Netherlands  |
| <b>UK Agent</b>                 | Dairy Ingredients Group - UK, Brent House, Brent, Tame Valley Industrial Estate, Tamworth Staffs. B77 5DF  |
| <b>Expert Technical Contact</b> | Jan H.P.M. Kerkhof (email jan.kerkhof@gist-brocades.infonet.com)   |
| <b>Recommended application</b>  | Diffusion test for detection of antibiotic and sulphonamide residues in raw commingled bovine milk.  |
| <b>Principle of test</b>        | The test detects inhibition of growth of bacteria. A dye is present in the growth medium which changes colour when bacterial growth occurs. If growth is inhibited, no colour change occurs. |
| <b>Cross-reactivity</b>         | Test detects a wide range of microbial inhibitors  |
| <b>Interference</b>             | Caution should be exercised to avoid contamination by antibiotics or other drugs, cleaning agents or disinfectants.  |
| <b>Apparatus supplied</b>       | Adhesive covers for plates   |
| <b>Extra apparatus required</b> | Water bath ; syringe (0.1 ml) with disposable tips   |
| <b>Reagents supplied</b>        | 96 well plates containing detection media  |
| <b>Extra reagents required</b>  | Negative control milk  |
| <b>Time (set-up)</b>            | 5 minutes  |
| <b>Time per test</b>            | <2½ hours, 96 samples including negative controls may be conveniently run at the same time   |
| <b>Operator skill level</b>     | Easy; some skill in pipetting required   |
| <b>Instructions</b>             | Concise and clear  |
| <b>Safety advice</b>            | No known hazards.  |
| <b>Reporting method</b>         | Positive or negative based on reference to colour chart. Samples with any trace of yellow colour were deemed negative i.e., only samples with entirely purple colour were positive           |
| <b>Ease of use</b>              | Requires minimal training (Ease of use = 5)  |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 0            |
| 1.0    | 0.9     | 30         | 0            | 10    | 9.1    | 30         | 0            |
| 2.0    | 1.9     | 30         | 30           | 15    | 13.6   | 30         | 30           |
| 3.0    | 2.8     | 30         | 30           | 20    | 18.2   | 30         | 30           |
| 4      | 3.8     | 30         | 30           | 30    | 27.2   | 30         | 30           |
|        |         |            |              | 50    | 45.4   | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic    | Sensitivity* ( $\mu\text{g kg}^{-1}$ ) |
|---------------|--|
| Penicillin G  | 2-4                                    |
| Sulphadiazine | 25-100                                 |

*Additional Comments*

None

*DDM/NV March 29, 1999 (Final revision)*

### Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Delvotest® P Mini  |
| <b>Supplier</b>                 | Gist-Brocades B.V. PO Box 1, 2600 MA Delft, The Netherlands  |
| <b>UK Agent</b>                 | Dairy Ingredients Group - UK, Brent House, Brent, Tame Valley Industrial Estate, Tamworth Staffs. B77 5DF  |
| <b>Expert Technical Contact</b> | Jan H.P.M. Kerkhof (email jan.kerkhof@gist-brocades.infonet.com)   |
| <b>Recommended application</b>  | Diffusion test for detection of antibiotic residues in raw bovine milk. (From individual animals and commingled milk.)   |
| <b>Principle of test</b>        | The test detects inhibition of growth of bacteria. A dye is present in the growth medium which changes colour when bacterial growth occurs. If growth is inhibited, no colour change occurs. |
| <b>Cross-reactivity</b>         | Test detects a wide range of microbial inhibitors  |
| <b>Interference</b>             | Caution should be exercised to avoid contamination by antibiotics or other drugs, cleaning agents or disinfectants. Samples that have soured should not be tested.                           |
| <b>Apparatus supplied</b>       | Tweezers; syringe (0.1 ml) with disposable tips  |
| <b>Extra apparatus required</b> | Delvo Incubator  |
| <b>Reagents supplied</b>        | Ampoules with <i>B.stearothermophilus</i> var. <i>calidolactis</i> in a solid agar medium; nutrient tablets.   |
| <b>Extra reagents required</b>  | Optional: Penicillinase  |
| <b>Time (set-up)</b>            | 10 minutes   |
| <b>Time per test</b>            | <2½ hours, 40 test samples may be conveniently run at the same time  |
| <b>Operator skill level</b>     | Easy; some skill in pipetting required   |
| <b>Instructions</b>             | Clear  |
| <b>Safety advice</b>            | No known hazards.  |
| <b>Reporting method</b>         | Positive or negative based on reference to colour chart. If any trace of yellow colour was detected, the sample was reported as negative i.e., completely purple.                            |
| <b>Ease of use</b>              | Requires minimal training (Ease of use = 5)  |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 0            |
| 1.0    | 0.9     | 30         | 0            | 20    | 18.2   | 30         | 0            |
| 2.0    | 1.9     | 30         | 0            | 30    | 27.2   | 30         | 0            |
| 3.0    | 2.8     | 30         | 30           | 40    | 36.3   | 30         | 30           |
| 4      | 3.8     | 30         | 30           | 50    | 45.4   | 30         | 30           |
| 5      | 4.5     | 30         | 30           | 70    | 63.6   | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

Minimum detection levels of DELVOTEST® P (ppb)

| Antibiotic          | Control time | 2½ hours   |
|---------------------|--------------|------------|
| Penicillin G        | 2.5          | 3          |
| Cloxacillin         | 15-25        | 25         |
| Dicloxacillin       | 15           | 20         |
| Oxacillin           | 10           | 10-15      |
| Nafcillin           | 5-10         | 10         |
| Ampicillin          | 3-4          | 5          |
| Amoxicillin         | 4            | 5-6        |
| Cephapirin          | 5            | 8-10       |
| Cefalonium          | 10-15        | 15-20      |
| Cefalexin           | 40-60        | 60-100     |
| Cefacetril          | 20-30        | 30-40      |
| Cefaperazon         | 40-60        | 80         |
| Tetracycline        | 50-100       | 200-300    |
| Oxytetracycline     | 50-200       | 300        |
| Chlotetracycline    | 50-100       | 250-300    |
| Tylosin             | 40-150       | 100-300    |
| Erythromicin        | 300-1000     | 500-1500   |
| Lincomycin          | 100-300      | 400-1000   |
| Spiramycin          | 1000-5000    | 5000+      |
| Gentamycin          | 200-800      | 1000-4000  |
| Neomycin            | 300-1000     | 1000-5000  |
| Dihydrostreptomycin | 2000-5000    | 10000+     |
| Kanamycin           | 7500+        | 10000+     |
| Chloramphenicol     | 2500-5000    | 5000-7500  |
| Sulfamethazine      | 10000+       | 10000+     |
| Sulfadimethoxine    | 10000+       | 10000+     |
| Sulfathiazole       | 10000+       | 10000+     |
| Sulfadiazine        | 10000+       | 10000+     |
| Dapsone             | 10000+       | 10000+     |
| Trimethoprim        | 10000+       | 10000+     |
|                     | 200-500      | 5000-15000 |

\* Control time = time at which blank control has just changed to yellow; based on yellow/purple colour

*Additional Comments*

None

DDM/NW March 29, 1999 (Final revision)

Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Delvotest® SP Mini   |
| <b>Supplier</b>                 | Gist-Brocades B.V. PO Box 1, 2600 MA Deift, The Netherlands  |
| <b>UK Agent</b>                 | Dairy Ingredients Group - UK, Brent House, Brent, Tame Valley Industrial Estate, Tamworth Staffs. B77 5DF  |
| <b>Expert Technical Contact</b> | Jan H.P.M. Kerkhof (email jan.kerkhof@gist-brocades.infonet.com)   |
| <b>Recommended application</b>  | Diffusion test for detection of antibiotic residues and sulphonamides in raw bovine milk. (From individual animals and commingled.)  |
| <b>Principle of test</b>        | The test detects inhibition of growth of bacteria. A dye is present in the growth medium which changes colour when bacterial growth occurs. If growth is inhibited, no colour change occurs. |
| <b>Cross-reactivity</b>         | Test detects a wide range of microbial inhibitors  |
| <b>Interference</b>             | Caution should be exercised to avoid contamination by antibiotics or other drugs, cleaning agents or disinfectants. Samples that have soured should not be tested.                           |
| <b>Apparatus supplied</b>       | Tweezers; syringe (0.1 ml) with disposable tips  |
| <b>Extra apparatus required</b> | Delvo Incubator  |
| <b>Reagents supplied</b>        | Ampoules with <i>B.stearothermophilus</i> var. <i>calidolactis</i> in a solid agar medium; nutrient tablets.   |
| <b>Extra reagents required</b>  | None   |
| <b>Time (set-up)</b>            | 10 minutes   |
| <b>Time per test</b>            | <3 hours, 40 test samples may be conveniently run at the same time   |
| <b>Operator skill level</b>     | Easy; some skill in pipetting required   |
| <b>Instructions</b>             | Clear  |
| <b>Safety advice</b>            | No known hazards.  |
| <b>Reporting method</b>         | Positive or negative based on reference to colour chart. Samples containing any yellow colour were deemed negative i.e., samples which were completely purple were positive                  |
| <b>Ease of use</b>              | Requires minimal training (Ease of use = 5)  |



*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 30         | 0            | 0     | 0      | 30         | 0            |
| 1.0    | 0.9     | 30         | 0            | 10    | 9.1    | 30         | 0            |
| 2.0    | 1.9     | 30         | 0            | 15    | 13.6   | 30         | 0            |
| 3.0    | 2.8     | 30         | 30           | 20    | 18.2   | 30         | 30           |
| 4      | 3.8     | 30         | 30           | 30    | 27.2   | 30         | 30           |
|        |         |            |              | 50    | 45.4   | 30         | 30           |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

Minimum detection levels of DELVOTEST® SP (ppb)

| Antibiotic          | Control time | 2½ hours  | 3 hours      |
|---------------------|--------------|-----------|--------------|
| Penicillin G        | 2            | 2.5       | 2.5          |
| Cloxacillin         | 15           | 15-25     | 20-25        |
| Dicloxacillin       | 10           | 10-15     | 10-15        |
| Oxacillin           | 5            | 10        | 10           |
| Nafcillin           | 5            | 5-8       | 10           |
| Ampicillin          | 2-3          | 3-4       | 3-5          |
| Amoxicillin         | 2            | 3-4       | 3-5          |
| Cephapirin          | 5            | 5-8       | 5-10         |
| Cefalonium          | 5-10         | 10-20     | 15-25        |
| Cefalexin           | 40-60        | 60-100    | 60-100       |
| Cefacetril          | 20           | 20-40     | 20-40        |
| Cefaperazon         | 40           | 60-100    | 60-100       |
| Tetracycline        | 100          | 200-400   | 300-600      |
| Oxytetracycline     | 100          | 200-400   | 40-50        |
| Chlotetracycline    | 100-150      | 200-400   | 300-600      |
| Tylosin             | 10-20        | 30-50     | 100          |
| Erythromicin        | 50           | 100-150   | 250          |
| Lincomycin          | 100          | 200       | 300-400      |
| Spiramycin          | 200          | 350-750   | 10000+       |
| Gentamycin          | 100-300      | 200-400   | 400-500      |
| Neomycin            | 100-200      | 300-1000  | 400-2000     |
| Dihydrostreptomycin | 300-500      | 1500-3000 | 2500-10000   |
| Kanamycin           | 2500         | 7500      | 10000+       |
| Chloramphenicol     | 2500         | 7500      | 75000-100000 |
| Sulfamethazine      | 25           | 50-100    | 100-200      |
| Sulfadimethoxine    | 50           | 50-100    | 100          |
| Sulfathiazole       | 50           | 50-100    | 100-150      |
| Sulfadiazine        | 50           | 50-100    | 100          |
| Dapsone             | 1            | 1-4       | 4-8          |
| Trimethoprim        | 50           | 100-300   | 500          |

\* Control time = time at which blank control has just changed to yellow

*Additional Comments*

None

DDM/NW March 29, 1999 (Final revision)

Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |   |
|---------------------------------|---|
| <b>Name:</b>                    | <b>Delvo-X-PRESS® L-II test for beta lactam residues</b>  |
| <b>Supplier</b>                 | Gist-Brocades B.V. PO Box 1, 2600 MA Delft, The Netherlands   |
| <b>UK Agent</b>                 | Dairy Ingredients Group - UK, Brent House, Brent, Tame Valley Industrial Estate, Tamworth Staffs. B77 5DF   |
| <b>Expert Technical Contact</b> | Jan H.P.M. Kerkhof (email jan.kerkhof@gist-brocades.infonet.com)  |
| <b>Recommended application</b>  | Rapid test for detection of beta lactam residues in commingled bovine milk. Milk samples can be frozen but should not be thawed more than twice.                                |
| <b>Principle of test</b>        | The test is a qualitative competitive receptor-enzyme assay .   |
| <b>Cross-reactivity</b>         | With other beta-lactams   |
| <b>Interference</b>             | Caution should be exercised to avoid contamination by antibiotics.  |
| <b>Apparatus supplied</b>       | None  |
| <b>Extra apparatus required</b> | ISR(P) workstation; micropipette (0.2 ml) plus disposable tips; tube racks; 250 ml polythene wash bottle; 5 ml polythene measuring beaker; 1 ml multipette plus disposable tips |
| <b>Reagents supplied</b>        | Reagent tubes, Tracer, Standard and diluent, Colour developer, wash solution.   |
| <b>Extra reagents required</b>  | Deionised or distilled water  |
| <b>Time (set-up)</b>            | 30 minutes  |
| <b>Time per test</b>            | 8 minutes (5 samples may be tested simultaneously)  |
| <b>Operator skill level</b>     | Modest, training required by a Gist-Brocades representative.  |
| <b>Instructions</b>             | Clear and concise   |
| <b>Safety advice</b>            | Detailed.   |
| <b>Reporting method</b>         | Positive or negative based instrument reading. samples which are negative have reader units < 00.   |
| <b>Ease of use</b>              | Requires some training (Ease of use = 10)   |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

Note: these results are provisional and should be regarded as indicative only

| Pen G* | Pen G** | No. trials | No. positive |
|--------|---------|------------|--------------|
| 0      | 0       | 50         | 2            |
| 3.0    | 2.8     | 20         | 13           |
| 4.0    | 3.8     | 20         | 18           |
| 5.0    | 4.5     | 20         | 19           |
| 6.0    | 5.4     | 20         | 20           |

\* Nominal concentration, \* Concentration as free base.

*Manufacturer's Sensitivity Declaration*

| Antibiotic   | Sensitivity* ( $\mu\text{g kg}^{-1}$ ) |
|--------------|--|
| Penicillin G | 2-4                                    |
| Ceftiofur    | 4-8                                    |
| Cephapirin   | 4-8                                    |
| Ampicillin   | 4-8                                    |
| Amoxicillin  | 4-8                                    |
| Cloxacillin  | 30-40                                  |

*Additional Comments*

Workstation displays instructions for each stage of test sequence.

*DDM/NW March 30, 1999 (Final revision)*

Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |   |
|---------------------------------|---|
| <b>Name:</b>                    | LacTek™ Penicillins Kit   |
| <b>Supplier</b>                 | IDEXX Laboratories Inc., One IDEXX Drive, Westbrook, ME 04092, USA  |
| <b>UK Agent</b>                 | Guildhay Ltd., 6 Riverside Business Centre, Walnut tree Close, Guildford, Surrey GU1 4UG  |
| <b>Expert Technical Contact</b> | Kalpna Raval (Kalpna.Raval@guildhay.co.uk)  |
| <b>Recommended application</b>  | Qualitative test for rapid detection of penicillins in raw (individual and comingled) and pasteurised milk.   |
| <b>Principle of test</b>        | The LacTek test is an enzyme-linked immunoabsorbant assay. Beta-lactams in milk compete with an enzyme tracer for antibody binding sites coated on an analysis tube. The presence of penicillins in the milk reduces the amount of tracer bound. When a reagent for colour development is introduced, the ultimate intensity of colour development is less than for a control sample. The colour measurement is carried out in a spectrophotometer. |
| <b>Cross-reactivity</b>         | The test cross reacts with amoxicillin, ampicillin, cloxacillin. And dicloxacillin  |
| <b>Interference</b>             | No known interference   |
| <b>Apparatus supplied</b>       | None  |
| <b>Extra apparatus required</b> | Timer, shaker, Reader/printer, micropipette (250 l) and disposable tips, dispensers (250 & 500 l), JetWasher.   |
| <b>Reagents supplied</b>        | Reaction tubes, Beta-lactam standard, diluent for standard, Tracer solution, diluent for tracer, colour developer, stop solution, wash concentrate, reconstitution labels, product insert.  |
| <b>Extra reagents required</b>  | De-ionised, distilled or RO water   |
| <b>Time (set-up)</b>            | Preparation of reagents, equilibration of test kit to room temperature.   |
| <b>Time per test</b>            | 10 minutes (5 samples may be tested at a time; time per test 2 minutes)   |
| <b>Operator skill level</b>     | Training is required by a LacTek representative   |
| <b>Instructions</b>             | Detailed and unambiguous.   |
| <b>Safety advice</b>            | No special precautions are required.  |
| <b>Reporting method</b>         | Reader/printer automatically displays and records results as positive or negative   |
| <b>Ease of use</b>              | Easy to use (rating = 11)   |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive | Clox* | Clox** | No. trials | No. positive |
|--------|---------|------------|--------------|-------|--------|------------|--------------|
| 0      | 0       | 20         | 0            | 0     | 0      | 20         | 0            |
| 2.0    | 1.9     | 20         | 0            | 5.0   | 4.5    | 20         | 1            |
| 2.5    | 2.3     | 20         | 1            | 7.5   | 6.8    | 20         | 20           |
| 3.0    | 2.8     | 20         | 6            | 10.0  | 9.1    | 20         | 20           |
| 3.5    | 3.3     | 20         | 20           | 12.5  | 11.4   | 20         | 20           |
| 5.0    | 4.7     | 20         | 20           | 15.0  | 13.6   | 20         | 20           |

\* Nominal concentration; \*\* Corrected to free base


*Manufacturer's Sensitivity Declaration*

| Antibiotic   | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|--------------|---------------------------------------|
| Penicillin G | At or below 4.8                       |
| Cloxacillin  | 90/95% sensitivity at 6.25            |


*Additional Comments*

Provision of dispensers facilitates liquid transfers and washing station is a helpful piece of ancillary equipment. In addition to the printout of the results, optical density readings are also provided which may be used to determine the sensitivity of the test. Test curves are shown below:

Lactek - response curve for cloxacillin



LACTEK - response curve for penicillin G



Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | <b>Penzym® 100</b>   |
| <b>Supplier</b>                 | UCB Bioproducts, Braine-L'Alleud, Belgium  |
| <b>UK Agent</b>                 | Axient, Axient Laboratories, Unit 20, Goldthorpe, Rotherham S63 9BL  |
| <b>Expert Technical Contact</b> | Jacques Degelaen ( <a href="mailto:jacques.degelaen@ucb-group.com">jacques.degelaen@ucb-group.com</a> )  |
| <b>Recommended application</b>  | Qualitative test for beta-lactam antibiotics, including penicillin G, ampicillin, amoxicillin, cloxacillin, cephalosporin C, cephalothin, cephalexin, cephuroxime, cephazolin, cephalosporin C, cephalothin, cephaloridine, cephadroxy and cefuroxime. Penzym 100 is designed for isolated or occasional use with milk e.g. on the milk tanker.  |
| <b>Principle of test</b>        | The Penzym test employs an enzyme called DD-carboxypeptidase. DD-carboxypeptidase specifically hydrolyses substrates of the R-D-alanine-D-alanine type with liberation of D-alanine. The alanine is detected by a coupled enzyme reaction that changes a dye from yellow to pink orange. Beta-lactam antibiotics specifically and quantitatively inhibit the enzyme. When they are present in milk the colour change does not occur. |
| <b>Cross-reactivity</b>         | The following drugs do not cross react at levels of 100ppb: sulfadiazine, sulfanilamide, sulfathiazole, sulfamethazine, sulfapyridine, sulfadimethoxine, tetracycline, oxytetracycline, chlortetracycline, doxycycline, gentamicin, neomycin, streptomycin, ivermectin, erythromycin, novobiocin, furosemide, trichlormethiazide, chlorothiazide, oxytocin, phenylbutazone, dexamethasone, dipyrone and dicloxacillin                |
| <b>Interference</b>             | Operator should not be under medical treatment by beta lactam antibiotics  |
| <b>Apparatus supplied</b>       | Tweezers, colour chart.  |
| <b>Extra apparatus required</b> | Dry incubator or water bath at 47°C, syringe and needle, micropipettes (10 & 50 µl), Eppendorf type vials (1.5 ml)   |
| <b>Reagents supplied</b>        | Enzyme vials, colour generating tablets  |
| <b>Extra reagents required</b>  | De-ionised water   |
| <b>Time (set-up)</b>            | Warm-up time of dry incubator or water bath  |
| <b>Time per test</b>            | Negative in 13 minutes, < 20 minutes for positive. (8 samples may be tested simultaneously.)   |
| <b>Operator skill level</b>     | Minimal  |
| <b>Instructions</b>             | Detailed and unambiguous.  |
| <b>Safety advice</b>            | No specific hazards  |
| <b>Reporting method</b>         | In 3 categories according to reference colour chart: yellow = positive,; between peach and peach yellow = positive , peach = caution, pink orange = negative   |
| <b>Ease of use</b>              | Relatively easy to use (rating = 10)   |

*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive<br>(caution) | Clox* | Clox** | No. trials | No.<br>positive |
|--------|---------|------------|---------------------------|-------|--------|------------|-----------------|
| 0      | 0       | 20         | 0                         | 0     | 0      | 20         | 0               |
| 2      | 1.9     | 20         | 0                         | 40    | 36     | 20         | (20)            |
| 3      | 2.8     | 20         | (20)                      | 60    | 54     | 20         | (20)            |
| 4      | 3.7     | 20         | (20)                      | 80    | 73     | 20         | 20              |
| 5      | 4.7     | 20         | (20)                      | 100   | 91     | 20         | 20              |
| 6      | 5.6     | 20         | 20                        |       |        |            |                 |
| 8      | 7.5     | 20         | 20                        |       |        |            |                 |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic      | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|-----------------|---------------------------------------|
| Penicillin G    | 4-6                                   |
| Amoxicillin     | 4-6                                   |
| Ampicillin      | 4-7                                   |
| Cephapirin      | 5-7                                   |
| Oxacillin       | 30-50                                 |
| Cloxacillin     | 60-100                                |
| Cephalexin      | 20-40                                 |
| Cephuroxime     | 50-100                                |
| Cephalonium     | 10-15                                 |
| Cephalosporin C | 8-10                                  |
| Cephalothin     | 4-6                                   |
| Cephaloridine   | 6-8                                   |
| Cephadroxyl     | 25-50                                 |
| Ceftiofur       | 40-70                                 |

*Additional Comments*

None

*DDM/NW March 12, 1999 (Final revision)*

Dairy Industry Federation Survey - Report on Antibiotic Test Kit

|                                 |  |
|---------------------------------|--|
| <b>Name:</b>                    | Penzym® 20   |
| <b>Supplier</b>                 | UCB Bioproducts, Braine-L'Alleud, Belgium  |
| <b>UK Agent</b>                 | Axient, Axient Laboratories, Unit 20, Goldthorpe, Rotherham S63 9BL  |
| <b>Expert Technical Contact</b> | Jacques Degelaen ( <a href="mailto:jacques.degelaen@ucb-group.com">jacques.degelaen@ucb-group.com</a> )  |
| <b>Recommended application</b>  | Qualitative test for beta-lactam antibiotics, including penicillin G, ampicillin, amoxicillin, cloxacillin, cephapirin, oxacillin, cephalixin, cephuroxime, cephalonium, cephalosporin C, cephalothin, cephaloridine, cephadroxy and ceftiofur. Penzym 20 is designed for isolated or occasional use with milk e.g. on the milk tanker.  |
| <b>Principle of test</b>        | The Penzym test employs an enzyme called DD-carboxypeptidase. DD-carboxypeptidase specifically hydrolyses substrates of the R-D-alanine-D-alanine type with liberation of D-alanine. The alanine is detected by a coupled enzyme reaction that changes a dye from yellow to pink orange. Beta-lactam antibiotics specifically and quantitatively inhibit the enzyme. When they are present in milk the colour change does not occur. |
| <b>Cross-reactivity</b>         | The following drugs do not cross react at levels of 100ppb: sulfadiazine, sulfanilamide, sulfathiazole, sulfamethazine, sulfapyridine, sulfadimethoxine, tetracycline, oxytetracycline, chlortetracycline, doxycycline, gentamicin, neomycin, streptomycin, ivermectin, erythromycin, novobiocin, furosemide, trichlormethiazide, chlorothiazide, oxytocin, phenylbutazone, dexamethasone, dipyrone and dicloxacillin                |
| <b>Interference</b>             | Operator should not be under medical treatment by beta lactam antibiotics  |
| <b>Apparatus supplied</b>       | Tweezers, syringe and 20 tips, colour chart.   |
| <b>Extra apparatus required</b> | Dry incubator or water bath at 47°C  |
| <b>Reagents supplied</b>        | Enzyme vials, colour generating tablets  |
| <b>Extra reagents required</b>  | None   |
| <b>Time (set-up)</b>            | Warm-up time of dry incubator or water bath  |
| <b>Time per test</b>            | Negative in 13 minutes, < 20 minutes for positive. (8 samples may be tested simultaneously.)   |
| <b>Operator skill level</b>     | Minimal  |
| <b>Instructions</b>             | Detailed and unambiguous.  |
| <b>Safety advice</b>            | No specific hazards  |
| <b>Reporting method</b>         | In 3 categories according to reference colour chart: yellow = positive,; between peach and peach yellow = positive , peach = caution, pink orange = negative   |
| <b>Ease of use</b>              | Very easy to use (rating = 6)  |



*Test results (antibiotic concentration in  $\mu\text{g kg}^{-1}$ )*

| Pen G* | Pen G** | No. trials | No. positive<br>(caution) | Clox* | Clox** | No. trials | No.<br>positive |
|--------|---------|------------|---------------------------|-------|--------|------------|-----------------|
| 0      | 0       | 20         | 0                         | 0     | 0      | 20         | 0               |
| 2      | 1.9     | 20         | 0                         | 40    | 36     | 20         | (20)            |
| 3      | 2.8     | 20         | (20)                      | 60    | 54     | 20         | (20)            |
| 4      | 3.7     | 20         | (20)                      | 80    | 73     | 20         | 20              |
| 5      | 4.7     | 20         | 20                        | 100   | 91     | 20         | 20              |
| 6      | 5.6     | 20         | 20                        |       |        |            |                 |
| 8      | 7.5     | 20         | 20                        |       |        |            |                 |

\* Nominal concentration; \*\* Corrected to free base

*Manufacturer's Sensitivity Declaration*

| Antibiotic      | Sensitivity ( $\mu\text{g kg}^{-1}$ ) |
|-----------------|---------------------------------------|
| Penicillin G    | 4-6                                   |
| Amoxicillin     | 4-6                                   |
| Ampicillin      | 4-7                                   |
| Cephapirin      | 5-7                                   |
| Oxacillin       | 30-50                                 |
| Cloxacillin     | 60-100                                |
| Cephalexin      | 20-40                                 |
| Cephuroxime     | 50-100                                |
| Cephalonium     | 10-15                                 |
| Cephalosporin C | 8-10                                  |
| Cephalothin     | 4-6                                   |
| Cephaloridine   | 6-8                                   |
| CephadroxyI     | 25-50                                 |
| Ceftiofur       | 40-70                                 |

*Additional Comments*

None

*DDMNV March 12, 1999 (Final revision)*